

Validation of Luminex xTAG Gastrointestinal Pathogen Panel with Stool

Specimens in S.T.A.R. Buffer

S.Y. Lee, Y. Elshimali, J. Lobos, and A. Scarsella

Consolidated Laboratory Services, Van Nuys, CA.



Shing-Yi Lee, Ph.D, HCLD
Consolidated Laboratory Services
7855 Haskell Avenue, Suite #302
Van Nuys, California 91406-1902
(800) 920-2022 • (818) 994-9714
E-mail: slee@cls-lab.com

Introduction

Gastrointestinal infections in both pediatric and adult patients account for significant morbidity and mortality worldwide. Diarrheal disease can be caused by a number of pathogens including viruses, bacteria, and parasites. For detection of enteric pathogen set, conventional diagnostic procedures involve culture, microscopy, and/or stool antigen tests. A newly FDA-approved test procedure, xTAG® Gastrointestinal Pathogen Panel (GPP) which includes 15 targets (4 pending approvals), overcomes most of those shortcomings and offers an economical and efficient management for clinical patients. This study evaluates and validates the analytical and clinical performance of the GPP assay.

Method

Two hundred and one (201) stool specimens were collected from 140 known (+) banked samples of Luminex®, 51 clinical patients, and 10 normal volunteers. For sample preparation, 1ml of S.T.A.R. (Stool Transport and Recovery) buffer with the internal control (MS2) was added to each tube (Abbott Master Mix Tube) containing stool-collected on the swab. After mixing thoroughly, incubate the tubes at 37°C water bath for 5 minutes. Store the tubes at -20°C until use. Thaw the tubes at 37°C water bath at least 5 minutes before centrifugation at 2000 rpm for 5 minutes. Remove any bubbles on the surface of the solution before testing procedure. Then, the samples were isolated with an open-mode protocol on Abbott's m2000 sp automatic instrument. The GPP kits (Luminex) were used following the manufacturer's instructions. The nucleic acid was amplified in a single multiplex RT-PCR reaction in a ProFlex thermocycler (Applied Biosystems), and the amplicon was hybridized with xMAP bead. Both the bead ID and the target reaction were detected by the Luminex²⁰⁰. The target-specific cut-off value was used with TDAS LSM software (xTAG® Data analysis Software LSM) to determine the presence/absence of the pathogens in the testing samples.

In addition, stored (-20°C) positive patient specimens were frozen and thawed for one to four times during a period between 2 weeks to five months to evaluate the effects of freeze/thaw cycles on the sample stability in STAR buffer.

Results

Each target was tested at least 6 times and all of them showed expected (+) results (Table 1 & 2). All normal stool samples were tested negative (data not shown). The external testing results were available for those 51 patients. Thirty-one out of fifty-one patients were tested negative both with GPP and outside lab (data not shown). The other 20 patient

Table 1: Accuracy and Reproducibility of Banked Specimens from Lunimex

Banked Samples	Signals (MFI)*	Banked Samples	Signals (MFI)*
Cryptosporidium	601	Rotavirus A	2532
Cryptosporidium	1286	Rotavirus A	2914
Cryptosporidium	591	Rotavirus A	2763
Cryptosporidium	886	Rotavirus A	4333
Cryptosporidium	1503	Rotavirus A	4736
Cryptosporidium	1575	Rotavirus A	3141
E.Coli. 0157	1819	C. difficile A/B	1960/2480
E.Coli. 0157	2005	C. difficile A/B	1492/1814
E.Coli. 0157	374	C. difficile A/B	1504/1548
E.Coli. 0157	2298	C. difficile A/B	1371/1831
E.Coli. 0157	1905	C. difficile A/B	1537/2102
E.Coli. 0157	2830	C. difficile A/B	2050/2008
STEC	1327/1301	Shigella	1721.5
STEC	1498/1339	Shigella	1803
STEC	653/252	Shigella	1771
STEC	1323/1462	Shigella	1770.5
STEC	1498/1339	Shigella	2185
STEC	1960/1407	Shigella	2028
ETEC LT/ST	1587.5/1741	Campylobacter	4787.5
ETEC LT/ST	2147/2351	Campylobacter	4000
ETEC LT/ST	3277/3423	Campylobacter	3392
ETEC LT/ST	3155.5/3472	Campylobacter	3546
ETEC LT/ST	3717/3939	Campylobacter	3801
ETEC LT/ST	3173/3389	Campylobacter	3985
Norovirus GI/GII	4528	Salmonella	1559/1420
Norovirus GI/GII	1494	Salmonella	2553/2322
Norovirus GI/GII	4426	Salmonella	4140.5/4032
Norovirus GI/GII	2136	Salmonella	4731/3967
Norovirus GI/GII	5109	Salmonella	4091/3776
Norovirus GI/GII	4780	Salmonella	4026/3598
Norovirus GI/GII	764	Salmonella	3615.5/3577
Giardia	925	Giardia	2288.5
Giardia	2282	Giardia	2169
Giardia	2304	Giardia	961

results were identified (+) targets either by GPP or outside lab (Table 3). Results of eight patients were found completely matched between two labs. There were 4 (+) results from GPP but were negative by outside lab. Another 6 (+) results picked up by GPP but they were not ordered by physicians. Two *E. histolytica dispar* (+) were not detected by GPP. A total of 17 positive samples were included in the freeze/thaw

Table 2. Accuracy and Reproducibility of 4 Pending FDA Approvals

Banked Samples	Signals (MFI)	Banked Samples	Signals (MFI)
Adenovirus 40-41	1609	Vibrio Cholerae	1662
Adenovirus 40-41	1579	Vibrio Cholerae	1622
Adenovirus 40-41	1677.5	Vibrio Cholerae	1377
Adenovirus 40-41	1603	Vibrio Cholerae	1333.5
Adenovirus 40-41	1540	Vibrio Cholerae	1378
Adenovirus 40-41	1700	Vibrio Cholerae	1370.5
Adenovirus 40-41	1578	Vibrio Cholerae	1318
Adenovirus 40-41	1629	Vibrio Cholerae	2062
Entamoeba Histolytica	1185	Yersinia Enterocolitica	2032.5
Entamoeba Histolytica	867	Yersinia Enterocolitica	1031
Entamoeba Histolytica	1137.5	Yersinia Enterocolitica	984
Entamoeba Histolytica	637.5	Yersinia Enterocolitica	1151
Entamoeba Histolytica	852.5	Yersinia Enterocolitica	1030
Entamoeba Histolytica	705.5	Yersinia Enterocolitica	836
Entamoeba Histolytica	701	Yersinia Enterocolitica	839
Entamoeba Histolytica	1046	Yersinia Enterocolitica	1203

Table 3. Comparison of Results between Luminex GPP and External Lab

Sample ID	GPP Results	External Reference Lab	Comments
140110705	Campylobacter	Campylobacter (+)	
140210498	Cryptosporidium	Cryptosporidium (+)	
140210152	Giardia	Giardia (+)	
140410383	Giardia	Giardia (+)	
140410830	Giardia	Giardia (+)	
140510710	Giardia	Giardia (+)	
140310952	Shigella	Shigella (+)	
140510685	Shigella	Shigella (+)	
140310273	Giardia		Giardia not ordered
140310427	ETEC/LTST		ETEC/LTST not ordered
140210628	Norovirus GI/GII		Norovirus not ordered
140410416	Norovirus GI/GII		Norovirus not ordered
140310531	Norovirus GI/GII		Norovirus not ordered
140310827	C. difficile A/B		C. difficile not ordered
140310952	Cryptosporidium		Cryptosporidium not ordered
140110705	Giardia	Giardia (-)	
140310531	Giardia	Giardia (-)	
140210718	C. difficile A/B	C. difficile A/B (-)	
140410693	ETEC/LTST	E. coli toxin (-)	
140410298	None	E. Histolytica-dispar Cysts (+)	
140410830	None	E. Histolytica-dispar Trophozoites (+)	

study (Table 4). All specimens were stable for at least two weeks with various freeze/thaw cycles, and some could be stable up to five months. However, targets with low signals (close to cut-off value, shown in **bold**) did not show reproducible results after two freeze/thaw cycles. No single reaction showed the inhibitory substance.

Table 4. Effect of Freeze/Thaw Cycles on the Stability of Samples in STAR Buffer

Sample ID	Results	F/T times	Storage Period
140110705	Campylobacter	4	five months
140210152	Giardia	1	one month
140210498	Cryptosporidium	3	four months
140210628	Norovirus GI/ GII	3	four months
140210718	C. difficile A/B	3	four months
140310273	Giardia	2	three months
140310427	ETEC/LTST	1	two weeks
140310827	C. difficile A/B	1	two weeks
N4	Giardia	1	two weeks
140410383	Giardia	1	one month
140410830	Giardia	1	three months
140510685	Shigella	2	one month
140510710	Giardia	1	one month
140610146	Norovirus GI/ GII	1	one month
140110705	Campylobacter, Giardia*	2	2 weeks
140310531	Giardia, Norovirus GI/GII*	2	2 weeks
140310952	Shigella, Cryptosporidium*	2	2 weeks

Conclusion

Our study results demonstrate that GPP is superior to the conventional methods in accuracy, sensitivity, and specificity. In addition, using S.T.A.R. buffer for the isolation of stool specimens instead of pre-treating stool with the bead as recommend by Luminex streamlines the testing procedure.

References

1. xTAG GPP Package Insert for use with Luminex 100/200, Rev A MLD-032-KPI-001 Rev. A; effective Date: January 2013
2. xPONENT® 3.1 Rev. 2 Software User Manual
3. TDAS GPP (US) User Manual
4. TDASLSM (US) User Manual